Merit Medical Systems, Inc.
Merit Hydrophilic Guide Wire
Special Premarket Notification 510(k)

JAN 16 2013

Section 5

510(k) Summary

General Provisions	Submitter Name:	Merit Medical Systems, Inc.
	Address:	1600 West Merit Parkway South Jordan, UT 84095
	Telephone Number:	(801) 208-4349
	Fax Number:	(801) 253-6967
	Contact Person:	Stephanie Erskine
	Registration Number:	1721504
	Correspondent Name:	Merit Medical Ireland Ltd.
	Address:	Parkmore Business Park Parkmore, Galway, Ireland
	Telephone Number:	(353) 91 703700 (3168)
	Fax Number:	(353) 91 771 888
	Contact Person:	Martha Folan
	Date of Preparation:	20-Dec-2012
	Registration Number:	9616662
Subject Device	Trade Name:	Merit Hydrophilic Guide Wire
	Common/Usual Name:	Hydrophilic Guide Wire
	Classification Name:	Catheter Guide Wire
Predicate Devices	Trade Name:	Merit Hydrophilic Guide Wire
	Classification Name:	Catheter Guide Wire
	Premarket Notification Predicate Device # 1 Merit Hydrophilic Guide Wire K120644 Manufacturer: Merit Medical Systems, Inc	
	Premarket Notification Predicate Device # 2: Radiofocus® Glidewire® K863138 Manufacturer: Terumo Medical Corp.	
Classification	Class II 21 CFR § 870.1330, Product code: DQX Division of Cardiovascular Devices	
Intended Use	The Merit Hydrophilic Guide Wire is intended to facilitate the placement of devices during diagnostic and interventional procedures.	

Merit Medical Systems, Inc.
Merit Hydrophilic Guide Wire
Special Premarket Notification 510(k)

Device Description	The Merit Hydrophilic Guide Wire consists of a jacketed core wire with a hydrophilic coating applied to the jacket. The wire will be offered in straight and angled configurations in various lengths.
Comparison to Predicate	Technological characteristics of the subject Merit Hydrophilic Guide Wire are substantially equivalent to those of the predicate, the Merit Hydrophilic Guide Wire [K120644]. The differences between the devices relates to the guide wire diameter size. The guide wire design remains unchanged. Predicate device #2 [K863138] Terumo Radiofocus® Glidewire® is the predicate for performance testing parameters.
Safety & Performance Tests	No performance standards have been established under section 514 of the Food, Drug and Cosmetic Act for these devices. A battery of testing was conducted in accordance with protocols based on requirements outlined in guidance's and industry standards and these were shown to meet the acceptance criteria that were determined to demonstrate substantial equivalence. Where appropriate, the tests were based on the requirements of the following documents: • FDA guidance Coronary and Cerebrovascular Guide Wire Guidance January 1995. • ISO 11070: 1998, *Sterile Single-Use Intravascular Catheter Introducers.* • ISO 11135-1: 2007 *Sterilization of health care products-Ethylene oxide- Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices.* • ASTM F1980-07 *Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices* • ISO 10993-1: 2009, *Biological Evaluation of Medical Devices Part 1: Evaluation and Testing within a risk management process*, and the FDA Modified ISO 10993 Test Profile FDA Memo G95-1. The Merit Hydrophilic Guide Wire was compared to the predicate device for various attributes that support substantial equivalence of the device. The differences in wire diameter between the modified device and the cleared device [K120644] has raised no new issues. Performance testing was based on Predicate device #2 [K863138] Terumo Radiofocus® Glidewire® . The following is a list of all the significant tests evaluated: Tensile Strength, Torque Strength, Torquability, Tip Flexibility, Coating Adherence/Integrity (including Evaluation using Anatomical Model), Catheter Compatibility (Durability), Surface, Fracture Test, Flex test, Size Designation/ Dimensions, Particulate Testing. As all test results were comparable to the predicate devices and all pre-determined acceptance criteria for tests were successful this has demonstrated the subject device is substantially equivalent to predicate devices.

Merit Medical Systems, Inc.
Merit Hydrophilic Guide Wire
Special Premarket Notification 510(k)

Summary of Substantial Equivalence	Based on the Indications for Use, design, safety and performance testing, the subject Merit Medical Hydrophilic Guide Wire meets the requirements that are considered essential for its intended use and is substantively equivalent to the predicate device, the Merit Hydrophilic Guide Wire manufactured by Merit Medical Systems Inc.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Merit Medical Systems, Inc.
Martha Folan
Senior Regulatory Affairs Specialist
Parkmore Business Park West
Parkmore, Galway, Ireland

JAN 16 2013

Re: K123609
Trade/Device Name: Merit Hydrophilic Guide Wire
Regulation Number: 21 CFR 870.1330
Regulation Name: Catheter Guide Wire
Regulatory Class: Class II
Product Code: DQX
Dated: November 19, 2012
Received: November 21, 2012

Dear Ms. Folan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Matthew G. Hillebrenner

for

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section 4

Indications for Use Statement

510 (k) Number: K123609

Device Name: Merit Hydrophilic Guide Wire

Indications for Use:

The Merit Hydrophilic Guide Wire is intended to facilitate the placement of devices during diagnostic and interventional procedures

Prescription Use X (Part 21 CFR 801 Subpart D)	AND/OR	Over-The-Counter Use ___ (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K123609